



12423 NE Whitaker Way
 Portland, OR 97230
 503-254-1794



Report Number: 23-012838/D002.R000
Report Date: 11/06/2023
ORELAP#: OR100028
Purchase Order:
Received: 10/30/23 10:32

Customer: Healthy TOKYO
Product identity: CBDTokyo 3.9(Batch 10-24-23)
Client/Metric ID: .
Laboratory ID: 23-012838-0001

Summary

Potency:

Analyte	Result (%)			
CBD	4.43		CBD-Total	4.43%
CBDV	0.00803		THC-Total	ND (Reported in percent of total sample)



Customer: Healthy TOKYO
1-22-2-1F Minami-Aoyama Minato-ku
1070062 TOKYO, Delaware
Japan

Product identity: CBDTokyo 3.9(Batch 10-24-23)

Client/Metric ID: .

Sample Date:

Laboratory ID: 23-012838-0001

Evidence of Cooling: No

Temp: 21.6

Relinquished by: shipping



Sample Results

Potency	Method: J AOAC 2015 V98-6 (mod) ^P		Units %	Batch: 2312440	Analyze: 11/2/23 3:40:00 AM
Analyte	As Received	Dry weight	LOQ	LOD	Notes
CBC	ND		0.0033	0.000686	
CBC-A	ND		0.0033	0.00108	
CBC-Total	ND		0.0061		
CBD	4.43		0.0327	0.0199	
CBD-A	ND		0.0033	0.000621	
CBD-Total	4.43		0.0355		
CBDV	0.00803		0.0033	0.00118	
CBDV-A	ND		0.0033	0.000490	
CBDV-Total	0.00803		0.0061		
CBE	ND		0.0033	0.000588	
CBG	ND		0.0033	0.00108	
CBG-A	ND		0.0033	0.000751	
CBG-Total	ND		0.0061		
CBL	ND		0.0033	0.00131	
CBL-A	ND		0.0033	0.000849	
CBL-Total	ND		0.0061		
CBN	ND		0.0033	0.000882	
CBT	ND		0.0033	0.000555	
Δ10-THC-9R	ND		0.0033	0.000817	
Δ10-THC-9S	ND		0.0033	0.000817	
Δ10-THC-Total	ND		0.0065		
Δ8-THC	ND		0.0033	0.00189	
Δ8-THCV	ND		0.0033	0.000849	
Δ9-THC	ND		0.0033	0.000817	
delta-9-THCP	ND		0.0033	0.00229	
exo-THC	ND		0.0033	0.000719	
THC-A	ND		0.0033	0.000751	
THC-Total	ND		0.0061		
THCV	ND		0.0033	0.00124	
THCV-A	ND		0.0033	0.000588	
THCV-Total	ND		0.0061		



● CBD
● CBDV



12423 NE Whitaker Way
Portland, OR 97230
503-254-1794



Report Number: 23-012838/D002.R000
Report Date: 11/06/2023
ORELAP#: OR100028
Purchase Order:
Received: 10/30/23 10:32

Potency	Method: J AOAC 2015 V98-6 (mod) ^p				Units %	Batch: 2312440	Analyze: 11/2/23 3:40:00 AM
Analyte	As Received	Dry weight	LOQ	LOD	Notes		
Total Cannabinoids	4.44						

Notes:

Samples analyzed for potency on a HPLC instrument.



12423 NE Whitaker Way
Portland, OR 97230
503-254-1794



Report Number: 23-012838/D002.R000
Report Date: 11/06/2023
ORELAP#: OR100028
Purchase Order:
Received: 10/30/23 10:32

Abbreviations

Limits: Action Levels per OAR-333-007-0400, OAR-333-007-0210, OAR-333-007-0220, CCR title 16-division 42. BCC-section 5723

Limit(s) of Quantitation (LOQ): The minimum levels, concentrations, or quantities of a target variable (e.g., target analyte) that can be reported with a specified degree of confidence.

Not Detected (ND): Analyte not detected to the specified limit.

^p = ISO/IEC 17025:2017 accredited method.

Units of Measure

% = Percentage of sample

% wt = µg/g divided by 10,000

Approved Signatory

Derrick Tanner
General Manager



12423 NE Whitaker Way
 Portland, OR 97230
 503-254-1794



Report Number: 23-012838/D002.R000
Report Date: 11/06/2023
ORELAP#: OR100028
Purchase Order:
Received: 10/30/23 10:32

Revision: 4 Document ID: 7148
 Legacy ID: Worksheet Validated 04/20/2021

Laboratory Quality Control Results

JAOAC2015 V98-6 Batch ID: 2312440

Laboratory Control Sample									
Analyte	LCS	Result	Spike	Units	% Rec	Limits		Evaluation	Notes
CBDVA	2	0.0313	0.0323	%	96.7	80.0	- 120	Acceptable	
CBDV	2	0.0333	0.0337	%	99.1	80.0	- 120	Acceptable	
CBE	2	0.0351	0.0358	%	97.9	80.0	- 120	Acceptable	
CBD	1	0.0317	0.0322	%	98.5	90.0	- 110	Acceptable	
CBD	1	0.0319	0.0329	%	96.9	80.0	- 120	Acceptable	
CBD	1	0.0363	0.0368	%	98.9	80.0	- 120	Acceptable	
CBD	1	0.0330	0.0313	%	105	90.0	- 110	Acceptable	
THCV	2	0.0332	0.0345	%	96.4	80.0	- 120	Acceptable	
38THCV	2	0.0288	0.0283	%	102	80.0	- 120	Acceptable	
THCV/A	2	0.0305	0.0312	%	98.0	80.0	- 120	Acceptable	
CBN	1	0.0321	0.0329	%	97.6	80.0	- 120	Acceptable	
exo-THC	2	0.0310	0.0315	%	98.5	80.0	- 120	Acceptable	
d9THC	1	0.0366	0.0365	%	100	90.0	- 110	Acceptable	
d8THC	1	0.0326	0.0340	%	95.9	90.0	- 110	Acceptable	
9S-d10THC	1	0.0326	0.0337	%	96.9	80.0	- 120	Acceptable	
CBL	2	0.0310	0.0332	%	93.5	80.0	- 120	Acceptable	
9R-d10THC	1	0.0327	0.0336	%	97.2	80.0	- 120	Acceptable	
CBC	2	0.0346	0.0342	%	101	80.0	- 120	Acceptable	
THCA	1	0.0322	0.0337	%	95.6	90.0	- 110	Acceptable	
CBCA	2	0.0326	0.0338	%	96.6	80.0	- 120	Acceptable	
CBLA	2	0.0341	0.0342	%	99.6	80.0	- 120	Acceptable	
38THCP	2	0.0306	0.0334	%	91.7	80.0	- 120	Acceptable	
CBT	2	0.0313	0.0343	%	91.4	80.0	- 120	Acceptable	

Method Blank						
Analyte	Result	LOQ	Units	Limits	Evaluation	Notes
CBDVA	<LOQ	0.00316	%	< 0.00316	Acceptable	
CBDV	<LOQ	0.00316	%	< 0.00316	Acceptable	
CBE	<LOQ	0.00316	%	< 0.00316	Acceptable	
CBD	<LOQ	0.00316	%	< 0.00316	Acceptable	
CBD	<LOQ	0.00316	%	< 0.00316	Acceptable	
THCV	<LOQ	0.00316	%	< 0.00316	Acceptable	
38THCV	<LOQ	0.00316	%	< 0.00316	Acceptable	
THCV/A	<LOQ	0.00316	%	< 0.00316	Acceptable	
CBN	<LOQ	0.00316	%	< 0.00316	Acceptable	
exo-THC	<LOQ	0.00316	%	< 0.00316	Acceptable	
d9THC	<LOQ	0.00316	%	< 0.00316	Acceptable	
d8THC	<LOQ	0.00316	%	< 0.00316	Acceptable	
9S-d10THC	<LOQ	0.00316	%	< 0.00316	Acceptable	
CBL	<LOQ	0.00316	%	< 0.00316	Acceptable	
9R-d10THC	<LOQ	0.00316	%	< 0.00316	Acceptable	
CBC	<LOQ	0.00316	%	< 0.00316	Acceptable	
THCA	<LOQ	0.00316	%	< 0.00316	Acceptable	
CBCA	<LOQ	0.00316	%	< 0.00316	Acceptable	
CBLA	<LOQ	0.00316	%	< 0.00316	Acceptable	
38THCP	<LOQ	0.00316	%	< 0.00316	Acceptable	
CBT	<LOQ	0.00316	%	< 0.00316	Acceptable	

Abbreviations
 ND - None Detected at or above MRI
 RPD - Relative Percent Difference
 LOQ - Limit of Quantitation

Units of Measure:
 %- Percent



12423 NE Whitaker Way
 Portland, OR 97230
 503-254-1794



Report Number: 23-012838/D002.R000
Report Date: 11/06/2023
ORELAP#: OR100028
Purchase Order:
Received: 10/30/23 10:32

Revision: 4 Document ID: 7148
 Legacy ID: Worksheet Validated 04/20/2021

Laboratory Quality Control Results

JAOAC2015 V98-6		Batch ID: 2312440						
Sample Duplicate		Sample ID: 23-012838-0001						
Analyte	Result	Org. Result	LOQ	Units	RPD	Limits	Evaluation	Notes
CBDVA	<LOQ	<LOQ	0.00328	%	NA	< 20	Acceptable	
CBDV	0.00804	0.00803	0.00328	%	0.157	< 20	Acceptable	
CBE	<LOQ	<LOQ	0.00328	%	NA	< 20	Acceptable	
CBD	<LOQ	<LOQ	0.00328	%	NA	< 20	Acceptable	
CBD	4.46	4.43	0.00328	%	0.539	< 20	Acceptable	
THCV	<LOQ	<LOQ	0.00328	%	NA	< 20	Acceptable	
Δ9THCV	<LOQ	<LOQ	0.00328	%	NA	< 20	Acceptable	
THCV/A	<LOQ	<LOQ	0.00328	%	NA	< 20	Acceptable	
CBN	<LOQ	<LOQ	0.00328	%	NA	< 20	Acceptable	
exo-THC	<LOQ	<LOQ	0.00328	%	NA	< 20	Acceptable	
Δ9THC	<LOQ	<LOQ	0.00328	%	NA	< 20	Acceptable	
Δ8THC	<LOQ	<LOQ	0.00328	%	NA	< 20	Acceptable	
9S-Δ10THC	<LOQ	<LOQ	0.00328	%	NA	< 20	Acceptable	
CBL	<LOQ	<LOQ	0.00328	%	NA	< 20	Acceptable	
9R-Δ10THC	<LOQ	<LOQ	0.00328	%	NA	< 20	Acceptable	
CBC	<LOQ	<LOQ	0.00328	%	NA	< 20	Acceptable	
THCA	<LOQ	<LOQ	0.00328	%	NA	< 20	Acceptable	
CBGA	<LOQ	<LOQ	0.00328	%	NA	< 20	Acceptable	
CBLA	<LOQ	<LOQ	0.00328	%	NA	< 20	Acceptable	
Δ9THCP	<LOQ	<LOQ	0.00328	%	NA	< 20	Acceptable	
CBT	<LOQ	<LOQ	0.00328	%	NA	< 20	Acceptable	

Abbreviations

- ND - None Detected at or above MRI
- RPD - Relative Percent Difference
- LOQ - Limit of Quantitation

Units of Measure:

%- Percent



12423 NE Whitaker Way
Portland, OR 97230
503-254-1794



Report Number: 23-012838/D002.R000
Report Date: 11/06/2023
ORELAP#: OR100028
Purchase Order:
Received: 10/30/23 10:32





Explanation of QC Flag Comments:

Code	Explanation
Q	Matrix interferences affecting spike or surrogate recoveries.
Q1	Quality control result biased high. Only non-detect samples reported.
Q2	Quality control outside QC limits. Data considered estimate.
Q3	Sample concentration greater than four times the amount spiked.
Q4	Non-homogenous sample matrix, affecting RPD result and/or % recoveries.
Q5	Spike results above calibration curve.
Q6	Quality control outside QC limits. Data acceptable based on remaining QC.
R	Relative percent difference (RPD) outside control limit.
R1	RPD non-calculable, as sample or duplicate results are less than five times the LOQ.
R2	Sample replicates RPD non-calculable, as only one replicate is within the analytical range.
LOQ1	Quantitation level raised due to low sample volume and/or dilution.
LOQ2	Quantitation level raised due to matrix interference.
B	Analyte detected in method blank, but not in associated samples.
B1	The sample concentration is greater than 5 times the blank concentration.
B2	The sample concentration is less than 5 times the blank concentration.